



# OIML CERTIFICATE OF CONFORMITY

To Xiamen Pinnacle Electrical Co.,Ltd.

This certificate attests the conformity of the type of the

Electronic Meterage Scale PB-615C ( Sale in domestic market ),

OL-615, ESP-615, TL-615 ( Export sales ) with the requirements

of the Recommendation R76(1992) of the

International Organization of Legal Metrology(OIML).

The issuing authority

Office of OIML Affairs

中国秘书处

December 7, 2004

The OIML member

王春平

December 7, 2004

Member State  
P.R.China

OIML Certificate No.  
R76/1992- CN1-04.06

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

Name: Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine  
Address: N0.9 Madiandonglu, Haidian District, Beijing, China

## Applicant

Name: Xiamen Pinnacle Electrical Co., Ltd.  
Address: 4F, Guangxia Building, North High-Tech Zone, Xiamen, Fujian, China

## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:  
Address:

## Identification of the certified type

Electronic Meterage Scale

Type	PB-615C (Sale in domestic market)	OL-615, ESP-615, TL-615 (Export sales)
Accuracy class	III	
Max (Max <sub>1</sub> /Max <sub>2</sub> )	6kg/15kg	
Min	40g	
c (c <sub>1</sub> /c <sub>2</sub> )	2g/5g	
n (n <sub>1</sub> /n <sub>2</sub> )	3000/3000	
Temperature range	0°C ~ 40°C	

The numbers of the associated test reports and the included pages:  
NIM-LXmw0402 35 pages each